



G.B. PANT UNIV.OF AGRIC.& TECH.
PANTNAGAR- 263 145 (UTTARAKHAND)

No. CA/PLP/6436

Dated: 24-05-2018

Corrigendum

Quotation Title: Quotation for Purchase of Orbital Incubator Shaker (BOD cum shaker) and Scientific Balance

Quotation Reference No.: CA/PLP/6309, Dated 24.04.2018

Orbital Incubator Shaker (BOD cum Shaker)	
Existing specifications (Previously published)	Proposed specifications (Now should be read as)
Temperature range 5 to 60°C, Shaking speed 20 – 250 rpm with digital indicator, DC Motor Power: 230V, 5A, Stainless Steel Platform with clamps of different size of flasks	Temperature range 5 to 60°C, Shaking speed 20 – 250 rpm with digital indicator, DC Motor Power: 230V, 5A, Stainless Steel Platform with clamps of different size of flasks Minimum Internal Volume: 200 Liters, Minimum shaking capacity: 1000ml × 12; Minimum shaking amplitude: 20mm

Last Date of Receiving Quotations : 08.06.2018

R.P.S.
24/5/2018

(R. P. Singh)
**Professor, Pl. Pathology
& Project Leader**