

G.B. PANT UNIVERSITY OF AGRI. & TECH. PANTNAGAR UTTARAKHAND
OFFICE OF THE HEAD DEPARTMENT OF MEDICINE
COLLEGE OF VETERINARY & ANIMAL SCIENCES PANTNAGAR

No. CVASc/VMD/Purchase/
M/s

Dated: March 27, 2018

Sub: Purchase of Double Beam Spectrophotometer

PLEASE CAREFULLY GO THROUGH THIS DOCUMENT AND ENSURE COMPLIANCE. NON-COMPLIANCE OF ANY CONDITION MAY MAKE YOUR OFFER INVALID

Dear Sirs,

We are interested in the purchase of the articles mentioned overleaf. Quotations are invited through speed post/registered post/courier/by hand from the firms to supply item. The details are also available on the University website. Please send your quotations in SEALED COVERS (sealed with sealing wax) so as to reach this office on or before 30.4.2018, while submitting quotations please note that:-

1. Rates quoted should be FOR Pantnagar including packing forwarding and transit insurance. The goods should be insured against theft, loss or breakage during transit by the supplier before dispatch is made. The responsibility in this respect will be of the supplying firm and not of the university.
2. The rate of Tax (GST) should be clearly indicated at which rate it is chargeable.
3. In case of the item being on DGS&D rate contract, please enclose a copy of the Rate contract.
4. Each quotation is to be enclosed in a cover. The cover should be sealed cover duly super-scribed, **“Quotation for the supply of Double Beam Spectrophotometer Your no. dated due date 30.4.2018”**. The cover should bear address of the undersigned.
5. The quotations shall be opened in the presence of representative of any or all the quoting firma **at 3.00 pm on next day or 2nd or 4th Saturday** at G.B. Pant University of Agriculture & Technology, Pantnagar.
6. Payment will be made by crossed cheque on the State Bank of India/UCO Bank/Punjab National Bank, Pantnagar or an account payee draft at the cost of supplier only after receipt of material in good condition.
7. The acceptance of quotations will rest with the Vice-chancellor who does not bind himself to accept the lowest rate quoted quotation and also reserves the right fully or partially any or all the quotations received without assigning any reason.
8. In case of dispute the matter shall be referred to the Vice-Chancellor of this University whose decision shall be binding to both the parties.
9. Please also quote your **GSTIN & Income Tax No.** on your quotation. Copy of **GST registration must be enclosed.**
10. In case of supply order being placed on you:
“As time is the essence of this order, the date of delivery should be strictly adhered to, otherwise the Vice-chancellor reserves the right not to accept delivery in part or full and claim the liquidated damages @ 1% per week subject to a maximum of 10% of total value of the supply order.
11. **You will have to deposit 3% of the total cost of item quoted by you as earnest money in the form of demand dfat or fixed deposit receipt or bankers cheque in favour of Comptroller G.B. Pant University of Agriculture & Technology, payable at SBI/PNB/UCO/Union Bank, Pantnagar.**
12. **The imported equipment above Rs. 3.00 lac may be purchased in Indian Currency provided Custom Duty Exemption Certificate is made available to the firm.**
13. On demand one sample of each quoted item may please be sent to judge the quality of material along with quotation. The sample should be received in this office on or before due date and opening of quotations.
14. The quotations are liable to be cancelled, if any of the above conditions are not complied with.
15. The validity of rates should be 180 days from the date of opening of quotations.

Yours faithfully

Signature & Seal of the Purchaser
Designation: Prof. & Head
Department: Veterinary Medicine

**DEPARTMENT OF EPIDEMIOLOGY & PREVENTIVE MEDICINE
COLLEGE OF VETERINARY AND ANIMAL SCIENCES PANTNAGAR**

Specifications:

Sl. No.	Instrument /Specification	Qty. Required	Rates Quoted
1.	<p>DOUBLE BEAM SPETROPHOTOMETER It should be Double Beam running on a minimum range of wave length of 200-1000 nm and having minimum wave length accuracy of minimum ± 0.3 nm. Wave length selection should be automatic with provision to select manually also. The detector should be silicon photovoltaic. There should be LED/LCD display.</p> <p>Light source should be tungsten and deuterium lamp. There should be Two USB ports (preferably one USB type C port) for attachment for Printer and Computer. It should be supplied with black and white laser printer with non-thermal paper input with a minimum of 8 mb on board memory.</p> <p>The minimum readability of the sample size should be 0.5 ml or less and the cuvette size should be 2 ml. It is to be supplied with extra one pair each of quartz and glass cuvettes.</p> <p>The equipment should have a warranty of three years.</p>	01 (One)	

Dated: 27.03.2018

Signature of Purchaser

Name: Dr. Mahesh Kumar
Designation: Prof. & Head

